




Potvrda o akreditaciji Accreditation Certificate

Ovime se utvrđuje da je
This is to recognize that

Varkom d.o.o.
Trg bana Jelačića 15, HR-42000 Varaždin
Laboratorij za otpadne vode
Ludbreška ulica 82/2, HR-42202 Trnovec

osposobljen prema zahtjevima norme
is competent according to
HRN EN ISO/IEC 17025:2017
(ISO/IEC 17025:2017;
EN ISO/IEC 17025:2017)
za/to carry out

Ispitivanje i uzorkovanje otpadnih voda
Testing and sampling of waste waters

u području opisanom u prilogu koji je sastavni dio ove potvrde o akreditaciji.
for the scope described in the annex which is the constituent part of this accreditation certificate.

Br./No.: 1539
Klasa/Ref.No.: 383-02/21-30/021
Urbroj/Id.No.: 569-02/7-25-24
Zagreb, 2025-05-21

Akreditacija istječe•Accreditation expiry: 2026-10-24
Prva akreditacija•Initial accreditation: 2016-10-12

HAA je potpisnica multilateralnog sporazuma s Europskom organizacijom za akreditaciju (EA)
HAA is a signatory of the European co-operation for Accreditation (EA) Multilateral Agreement

Ravnateljica:
Director General:
mr. sc. Mirela Zečević


HAA-Ob-7/7-1/izdanje/Issue 6



HAA

Hrvatska akreditacijska agencija
Croatian Accreditation Agency

PRILOG POTVRDI O AKREDITACIJI br: 1539*Annex to Accreditation Certificate Number:*

Klasa/Ref. No.: 383-02/21-30/021

Urbroj/Id. No.: 569-02/7-25-23

Datum izdanja priloga /Annex Issued on: 2025-05-21

Zamjenjuje prilog/Replaces Annex:

Klasa/Ref. No.: 383-02/21-30/021

Urbroj/Id. No.: 569-02/7-22-38

Datum /Date: 2022-12-16

Norma: HRN EN ISO/IEC 17025:2017*Standard: (ISO/IEC 17025:2017; EN ISO/IEC 17025:2017)***Akreditacija istječe: 2026-10-24***Accreditation expiry:***Prva akreditacija: 2016-10-12***Initial accreditation:***Akreditirani laboratorij***Accredited Laboratory***Varkom d.o.o.**

Trg bana Jelačića 15, HR-42000 Varaždin

Laboratorij za otpadne vode

Ludbreška ulica 82/2, HR-42202 Trnovec

Područje akreditacije:*Scope of Accreditation:***Ispitivanje i uzorkovanje otpadnih voda***Testing and sampling of waste waters*Važeće izdanje Priloga dostupno je na web adresi: www.akreditacija.hr
*Valid issue of the Annex is available at the web address: www.akreditacija.hr***Ravnateljica:***Director General:*42. 
mr. sc. Mirela Zečević

PODRUČJE AKREDITACIJE / SCOPE OF ACCREDITATION

Laboratorij za otpadne vode
Ludbreška ulica 82/2, HR-42 202 Trnovec

Br. No.	Materijali/Proizvodi Materials/Products	Vrsta ispitivanja/Svojstvo Type of test/Property Raspon/Range	Metoda ispitivanja Test method
1.	Otpadna voda Waste water	Određivanje suspendiranih tvari – Metoda filtriranjem kroz filter od staklenih vlakana <i>Determination of suspended Solids – Method by filtration through glass fibre filters</i> Granica kvantifikacije/ <i>Limit of quantification</i> 2 mg/l	HRN EN 872:2008 (EN 872:2005)
2.		Određivanje indeksa kemijske potrošnje kisika – Metoda s malim zatvorenim epruvetama <i>Determination of the chemical oxygen consumption index – Method with small closed tubes</i> (51 -2000) mg O ₂ /l	Vlastita metoda <i>In-house method</i> Q-RUL2-06, Točka / Clause: 5.4 Izdanje/revizija <i>Edition/revision</i> 2/1 2022-09-28 modificirana <i>modified</i> HRN ISO 15705:2003 (ISO 15705:2002)
3.		Određivanje indeksa kemijske potrošnje kisika – Metoda s malim zatvorenim epruvetama <i>Determination of the chemical oxygen consumption index – Method with small closed tubes</i> (15 – 2000) mg O ₂ /l	Vlastita metoda <i>In-house method</i> Q-RUL2-06, Točka / Clause: 5.5 Izdanje/revizija <i>Edition/revision</i> 2/1 2022-09-28 (Kivetni test/ Cuvette test HACH LANGE, LCK 314 i LCK 514)
4.		Određivanje biokemijske potrošnje kisika nakon 5 dana (BPK ₅) <i>Determination of biochemical oxygen demand after 5 days (BOD₅)</i> (10-1500) mg O ₂ /l	HRN EN ISO 5815-1:2019 (ISO 5815-1:2019, EN ISO 5815-1:2019)

**HAA**Hrvatska akreditacijska agencija
Croatian Accreditation AgencyPrilog potvrdi o akreditaciji
Annex to the Accreditation Certificate
Datum izdanja priloga/ Annex Issued on
Zamjenjuje prilog od/Replaces Annex dated:

1539

2025-05-21

2022-12-16

Br. No.	Materijali/Proizvodi Materials/Products	Vrsta ispitivanja/Svojstvo Type of test/Property Raspon/Range	Metoda ispitivanja Test method
5.	Otpadna voda Waste water	Određivanje pH Determination of pH 4-9 pH	HRN EN ISO 10523:2012 (ISO 10523:2008; EN ISO 10523:2012)
6.	Otpadna voda Waste water	Uzorkovanje	HRN ISO 5667-10:2020 (ISO 5667-10:2020)

